

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050223\
 Data File : VV030976.D
 Acq On : 02 May 2023 12:28
 Operator : SY/MD
 Sample : VV0502WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK210

Quant Time: May 03 06:29:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM050223.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 03 06:25:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	4116	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	3841	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2245	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.281	65	1578	0.452	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	90.000%
4) 1,2-Dichloroethane-d4	4.970	65	1497	0.503	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.000%
7) 1,2-Dichloropropane-d6	5.996	67	1333	0.524	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	104.000%
8) Toluene-d8	7.254	98	4656	0.500	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	1042	0.522	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	104.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050223\
Data File : VV030976.D
Acq On : 02 May 2023 12:28
Operator : SY/MD
Sample : VV0502WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK210

Quant Time: May 03 06:29:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM050223.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 03 06:25:35 2023
Response via : Initial Calibration

