

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050223\
 Data File : VV030978.D
 Acq On : 02 May 2023 15:11
 Operator : SY/MD
 Sample : VSTDCCC0.5EC
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5295

Quant Time: May 03 06:29:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM050223.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 03 06:25:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	4097	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	3809	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2157	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	1671	0.480	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	96.000%		
4) 1,2-Dichloroethane-d4	4.946	65	1457	0.492	ug/L	-0.02
Spiked Amount 0.500	Range 70 - 130		Recovery =	98.000%		
7) 1,2-Dichloropropane-d6	5.995	67	1313	0.520	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	104.000%		
8) Toluene-d8	7.253	98	4665	0.505	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	102.000%		
10) 1,1,2,2-Tetrachloroeth...	10.146	84	1013	0.512	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	102.000%		
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.280	62	1549	0.489	ug/L	99
6) Trichloroethene	5.852	95	1450	0.484	ug/L	99
9) 1,2-Dibromoethane	8.290	107	900	0.524	ug/L	97
12) 1,2,3-Trichloropropane	10.214	75	719	0.542	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.357	75	159	0.517	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050223\
Data File : VV030978.D
Acq On : 02 May 2023 15:11
Operator : SY/MD
Sample : VSTDCCC0.5EC
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5295

Quant Time: May 03 06:29:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM050223.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 03 06:25:35 2023
Response via : Initial Calibration

