

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050223\
 Data File : VV030973.D
 Acq On : 02 May 2023 10:57
 Operator : SY/MD
 Sample : VSTD00129
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD001229

Quant Time: May 03 06:22:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM050223.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 03 06:21:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	4172	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	3956	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2269	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	3387	0.970	ug/L	0.00
4) 1,2-Dichloroethane-d4	4.946	65	3198	1.090	ug/L	-0.02
7) 1,2-Dichloropropane-d6	5.995	67	2789	1.127	ug/L	0.00
8) Toluene-d8	7.254	98	10154	1.046	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.146	84	2237	1.058	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.280	62	3126	0.944	ug/L	98
6) Trichloroethene	5.852	95	3001	0.844	ug/L	99
9) 1,2-Dibromoethane	8.290	107	1941	1.012	ug/L #	95
12) 1,2,3-Trichloropropane	10.214	75	1556	1.070	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.357	75	350	0.981	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050223\
Data File : VV030973.D
Acq On : 02 May 2023 10:57
Operator : SY/MD
Sample : VSTD00129
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD001229

Quant Time: May 03 06:22:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM050223.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 03 06:21:13 2023
Response via : Initial Calibration

