

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050223\
 Data File : VV030974.D
 Acq On : 02 May 2023 11:22
 Operator : SY/MD
 Sample : VSTD00230
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD002230

Quant Time: May 03 06:22:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM050223.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 03 06:21:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	4202	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	3984	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2246	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	6302	1.792	ug/L	0.00
4) 1,2-Dichloroethane-d4	4.946	65	5762	1.950	ug/L	-0.02
7) 1,2-Dichloropropane-d6	5.995	67	5075	2.036	ug/L	0.00
8) Toluene-d8	7.253	98	18787	1.922	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.146	84	4078	1.914	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.287	62	6142	1.842	ug/L	98
6) Trichloroethene	5.852	95	5926	1.654	ug/L	98
9) 1,2-Dibromoethane	8.290	107	3739	1.936	ug/L #	94
12) 1,2,3-Trichloropropane	10.214	75	2995	2.080	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.356	75	685	1.940	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050223\
Data File : VV030974.D
Acq On : 02 May 2023 11:22
Operator : SY/MD
Sample : VSTD00230
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD002230

Quant Time: May 03 06:22:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM050223.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 03 06:21:13 2023
Response via : Initial Calibration

