

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050321\
 Data File : VV021248.D
 Acq On : 03 May 2021 16:14
 Operator : SY/MD
 Sample : M2209-11
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG7C9

Quant Time: May 04 05:26:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM050321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon May 03 18:54:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	708514	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	680104	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.255	152	337659	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	146382	46.58	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	93.16%		
7) Chloroethane-d5	1.568	69	137077	48.81	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	97.62%		
11) 1,1-Dichloroethene-d2	2.108	63	291534	33.25	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	66.50%		
21) 2-Butanone-d5	3.896	46	231091	91.63	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	91.63%		
24) Chloroform-d	4.352	84	463228	46.31	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	92.62%		
26) 1,2-Dichloroethane-d4	5.037	65	366793	48.98	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	97.96%		
32) Benzene-d6	5.053	84	772874	49.34	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	98.68%		
36) 1,2-Dichloropropane-d6	6.072	67	196277	48.78	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	97.56%		
41) Toluene-d8	7.320	98	779040	48.00	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	96.00%		
43) trans-1,3-Dichloroprop...	7.625	79	139072	46.47	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	92.94%		
47) 2-Hexanone-d5	8.092	63	154797	83.31	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	83.31%		
56) 1,1,2,2-Tetrachloroeth...	10.223	84	276099	42.88	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	85.76%		
66) 1,2-Dichlorobenzene-d4	11.632	152	342626	51.49	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	102.98%		
Target Compounds						Qvalue
13) Acetone	2.188	43	11304	3.79	ug/L	97
14) Carbon disulfide	2.294	76	18221	1.64	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050321\
Data File : VV021248.D
Acq On : 03 May 2021 16:14
Operator : SY/MD
Sample : M2209-11
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG7C9

Quant Time: May 04 05:26:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM050321WMA.M
Quant Title : VOC Analysis
QLast Update : Mon May 03 18:54:51 2021
Response via : Initial Calibration

