

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW050321\
 Data File : VW021242.D
 Acq On : 03 May 2021 13:52
 Operator : SY/MD
 Sample : M2249-02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG7G9

Quant Time: May 04 05:24:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM050321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon May 03 18:54:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	594132	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	572867	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.255	152	285478	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	167134	63.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	126.84%
7) Chloroethane-d5	1.571	69	150666	63.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	127.94%
11) 1,1-Dichloroethene-d2	2.108	63	321137	43.68	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.36%
21) 2-Butanone-d5	3.896	46	244468	115.60	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	115.60%
24) Chloroform-d	4.352	84	501054	59.73	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	119.46%
26) 1,2-Dichloroethane-d4	5.037	65	394849	62.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	125.76%#
32) Benzene-d6	5.053	84	847065	64.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	128.40%#
36) 1,2-Dichloropropane-d6	6.072	67	216238	63.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	127.62%#
41) Toluene-d8	7.320	98	850084	62.18	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	124.36%#
43) trans-1,3-Dichloroprop...	7.625	79	148925	59.08	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	118.16%
47) 2-Hexanone-d5	8.092	63	172609	110.29	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	110.29%
56) 1,1,2,2-Tetrachloroeth...	10.223	84	303710	56.00	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	112.00%
66) 1,2-Dichlorobenzene-d4	11.632	152	368049	65.42	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	130.84%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050321\
Data File : VV021242.D
Acq On : 03 May 2021 13:52
Operator : SY/MD
Sample : M2249-02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG7G9

Quant Time: May 04 05:24:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM050321WMA.M
Quant Title : VOC Analysis
QLast Update : Mon May 03 18:54:51 2021
Response via : Initial Calibration

