

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050321\
 Data File : VV021252.D
 Acq On : 03 May 2021 17:49
 Operator : SY/MD
 Sample : M2177-10ME
 Misc : 2.97g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG576ME

Quant Time: May 04 05:26:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM050321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon May 03 18:54:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	646876	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.860	117	645545	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.259	152	353086	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	214051	74.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	149.20%#
7) Chloroethane-d5	1.568	69	136949	53.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	106.82%
11) 1,1-Dichloroethene-d2	2.089	63	478501	59.78	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	119.56%
21) 2-Butanone-d5	3.912	46	282683	122.77	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	122.77%
24) Chloroform-d	4.349	84	680392	74.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	148.98%#
26) 1,2-Dichloroethane-d4	5.034	65	528691	77.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	154.66%#
32) Benzene-d6	5.047	84	1151915	77.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	154.96%#
36) 1,2-Dichloropropane-d6	6.072	67	295131	77.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	154.56%#
41) Toluene-d8	7.317	98	1163927	75.55	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	151.10%#
43) trans-1,3-Dichloroprop...	7.625	79	210191	74.00	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	148.00%#
47) 2-Hexanone-d5	8.121	63	262403	148.79	ug/L	0.03
Spiked Amount	100.000	Range	45 - 130	Recovery	=	148.79%#
56) 1,1,2,2-Tetrachloroeth...	10.230	84	446204	73.01	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	146.02%#
66) 1,2-Dichlorobenzene-d4	11.635	152	530628	76.26	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	152.52%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050321\
Data File : VV021252.D
Acq On : 03 May 2021 17:49
Operator : SY/MD
Sample : M2177-10ME
Misc : 2.97g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG576ME

Quant Time: May 04 05:26:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM050321WMA.M
Quant Title : VOC Analysis
QLast Update : Mon May 03 18:54:51 2021
Response via : Initial Calibration

