

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW050322\
 Data File : VW025803.D
 Acq On : 03 May 2022 19:13
 Operator : SY/MD
 Sample : N2615-15
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH0B1

Quant Time: May 04 05:09:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR050322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 04 05:06:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	153017	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	156465	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	73371	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.311	65	77465	4.619	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.400%
7) Chloroethane-d5	1.571	69	69954	4.742	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.800%
11) 1,1-Dichloroethene-d2	2.111	63	107660	3.379	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.600%
20) 2-Butanone-d5	3.899	46	92000	40.081	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	80.160%
24) Chloroform-d	4.352	84	109664	4.365	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.400%
26) 1,2-Dichloroethane-d4	5.037	65	50129	4.509	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.200%
32) Benzene-d6	5.053	84	225315	4.484	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.600%
36) 1,2-Dichloropropane-d6	6.069	67	65274	4.417	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.400%
41) Toluene-d8	7.317	98	188193	4.140	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.800%
43) trans-1,3-Dichloroprop...	7.625	79	16157	3.454	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	69.000%
46) 2-Hexanone-d5	8.088	63	71127	39.012	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	78.020%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	39121	4.193	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.800%
66) 1,2-Dichlorobenzene-d4	11.622	152	59796	4.646	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050322\
Data File : VV025803.D
Acq On : 03 May 2022 19:13
Operator : SY/MD
Sample : N2615-15
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH0B1

Quant Time: May 04 05:09:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR050322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 04 05:06:02 2022
Response via : Initial Calibration

