

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV050322\  
 Data File : VV025819.D  
 Acq On : 04 May 2022 02:16  
 Operator : SY/MD  
 Sample : VV0503WBL02  
 Misc : 25mL/MSVOA\_V/WATER  
 ALS Vial : 42 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VBLK356

Quant Time: May 04 07:58:12 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR050322WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Wed May 04 07:57:41 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min)  |
|-------------------------------|--------|-------|----------|----------|-------|-----------|
| -----                         |        |       |          |          |       |           |
| Internal Standards            |        |       |          |          |       |           |
| 1) 1,4-Difluorobenzene        | 5.619  | 114   | 121235   | 5.000    | ug/L  | 0.00      |
| 28) Chlorobenzene-d5          | 8.854  | 117   | 122542   | 5.000    | ug/L  | 0.00      |
| 58) 1,4-Dichlorobenzene-d4    | 11.249 | 152   | 52798    | 5.000    | ug/L  | 0.00      |
| System Monitoring Compounds   |        |       |          |          |       |           |
| 4) Vinyl Chloride-d3          | 1.310  | 65    | 71458    | 5.377    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =     | 107.600%  |
| 7) Chloroethane-d5            | 1.571  | 69    | 67215    | 5.750    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =     | 115.000%  |
| 11) 1,1-Dichloroethene-d2     | 2.111  | 63    | 101137   | 4.006    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =     | 80.200%   |
| 20) 2-Butanone-d5             | 3.899  | 46    | 117504   | 64.613   | ug/L  | 0.00      |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =     | 129.220%  |
| 24) Chloroform-d              | 4.352  | 84    | 108122   | 5.432    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 108.600%  |
| 26) 1,2-Dichloroethane-d4     | 5.037  | 65    | 52642    | 5.976    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 119.600%  |
| 32) Benzene-d6                | 5.053  | 84    | 213452   | 5.424    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 108.400%  |
| 36) 1,2-Dichloropropane-d6    | 6.072  | 67    | 67331    | 5.817    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =     | 116.400%  |
| 41) Toluene-d8                | 7.317  | 98    | 168622   | 4.736    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 94.800%   |
| 43) trans-1,3-Dichloroprop... | 7.625  | 79    | 19084    | 5.209    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =     | 104.200%  |
| 46) 2-Hexanone-d5             | 8.088  | 63    | 88076    | 61.682   | ug/L  | 0.00      |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =     | 123.360%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.217 | 84    | 42346    | 5.795    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =     | 115.800%  |
| 66) 1,2-Dichlorobenzene-d4    | 11.625 | 152   | 54591    | 5.895    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =     | 117.800%  |
| Target Compounds              |        |       |          |          |       |           |
| 16) Methylene chloride        | 2.510  | 84    | 3208     | 0.283    | ug/L  | Qvalue 91 |
| -----                         |        |       |          |          |       |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV050322\  
Data File : VV025819.D  
Acq On : 04 May 2022 02:16  
Operator : SY/MD  
Sample : VV0503WBL02  
Misc : 25mL/MSVOA\_V/WATER  
ALS Vial : 42 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VBLK356

Quant Time: May 04 07:58:12 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR050322WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Wed May 04 07:57:41 2022  
Response via : Initial Calibration

