

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW050322\
 Data File : VW025835.D
 Acq On : 04 May 2022 08:34
 Operator : SY/MD
 Sample : N2678-11
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH0E2

Quant Time: May 04 08:56:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR050322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 04 07:57:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	138829	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	141929	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	66864	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	67324	4.424	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.400%
7) Chloroethane-d5	1.571	69	66663	4.980	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.600%
11) 1,1-Dichloroethene-d2	2.111	63	99895	3.456	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.200%
20) 2-Butanone-d5	3.896	46	115184	55.310	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.620%
24) Chloroform-d	4.352	84	111022	4.871	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.400%
26) 1,2-Dichloroethane-d4	5.034	65	50965	5.052	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.000%
32) Benzene-d6	5.053	84	216690	4.754	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.000%
36) 1,2-Dichloropropane-d6	6.069	67	66683	4.974	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.400%
41) Toluene-d8	7.317	98	170946	4.146	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.000%
43) trans-1,3-Dichloroprop...	7.625	79	17212	4.056	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.200%
46) 2-Hexanone-d5	8.088	63	85380	51.626	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.260%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	44737	5.286	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.800%
66) 1,2-Dichlorobenzene-d4	11.622	152	61950	5.282	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.600%
Target Compounds						
					Qvalue	
13) Acetone	2.185	43	19648	14.335	ug/L	93
16) Methylene chloride	2.513	84	2584	0.199	ug/L	89
21) 2-Butanone	3.998	43	19979	10.894	ug/L	84
33) Benzene	5.108	78	11239	0.236	ug/L	100
42) Toluene	7.394	91	16441	0.325	ug/L	94
53) m,p-Xylene	9.149	106	1345	0.065	ug/L	83
54) o-Xylene	9.548	106	1609	0.083	ug/L	80
63) 1,2,4-Trimethylbenzene	10.921	105	3076	0.087	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050322\
Data File : VV025835.D
Acq On : 04 May 2022 08:34
Operator : SY/MD
Sample : N2678-11
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 58 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH0E2

Quant Time: May 04 08:56:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR050322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 04 07:57:41 2022
Response via : Initial Calibration

