

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW050322\
 Data File : VW025798.D
 Acq On : 03 May 2022 17:15
 Operator : SY/MD
 Sample : N2615-10
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH0A6

Quant Time: May 04 05:08:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR050322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 04 05:06:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	153641	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	160582	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	72793	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	79610	4.727	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.600%
7) Chloroethane-d5	1.571	69	69899	4.719	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.400%
11) 1,1-Dichloroethene-d2	2.111	63	109988	3.438	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.800%
20) 2-Butanone-d5	3.896	46	93426	40.537	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.080%
24) Chloroform-d	4.352	84	111491	4.420	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.400%
26) 1,2-Dichloroethane-d4	5.034	65	50311	4.507	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.200%
32) Benzene-d6	5.053	84	226796	4.398	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.000%
36) 1,2-Dichloropropane-d6	6.072	67	66264	4.369	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.400%
41) Toluene-d8	7.317	98	185466	3.975	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.600%
43) trans-1,3-Dichloroprop...	7.628	79	17211	3.585	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	71.600%
46) 2-Hexanone-d5	8.088	63	71401	38.158	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	76.320%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	37862	3.954	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	58930	4.615	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050322\
Data File : VV025798.D
Acq On : 03 May 2022 17:15
Operator : SY/MD
Sample : N2615-10
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH0A6

Quant Time: May 04 05:08:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR050322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 04 05:06:02 2022
Response via : Initial Calibration

