

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW050322\
 Data File : VW025800.D
 Acq On : 03 May 2022 18:02
 Operator : SY/MD
 Sample : N2615-12
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH0A8

Quant Time: May 04 05:08:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR050322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 04 05:06:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	157973	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	161094	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	76149	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	78104	4.511	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.200%
7) Chloroethane-d5	1.571	69	68545	4.500	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.000%
11) 1,1-Dichloroethene-d2	2.111	63	111128	3.378	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.600%
20) 2-Butanone-d5	3.899	46	96642	40.783	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.560%
24) Chloroform-d	4.352	84	108989	4.202	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.000%
26) 1,2-Dichloroethane-d4	5.037	65	49712	4.331	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.600%
32) Benzene-d6	5.053	84	229376	4.434	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.600%
36) 1,2-Dichloropropane-d6	6.069	67	66082	4.343	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.800%
41) Toluene-d8	7.316	98	187831	4.013	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.200%
43) trans-1,3-Dichloroprop...	7.625	79	16982	3.526	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	70.600%
46) 2-Hexanone-d5	8.088	63	71574	38.129	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	76.260%
56) 1,1,2,2-Tetrachloroeth...	10.213	84	38212	3.978	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.600%
66) 1,2-Dichlorobenzene-d4	11.622	152	59399	4.447	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050322\
Data File : VV025800.D
Acq On : 03 May 2022 18:02
Operator : SY/MD
Sample : N2615-12
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH0A8

Quant Time: May 04 05:08:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR050322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 04 05:06:02 2022
Response via : Initial Calibration

