

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050323\
 Data File : VV030981.D
 Acq On : 03 May 2023 10:53
 Operator : SY/MD
 Sample : VV0503WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK332

Quant Time: May 04 05:37:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu May 04 05:37:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	121264	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	118564	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	60868	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	44991	4.503	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.000%
7) Chloroethane-d5	1.539	69	38639	4.741	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.800%
11) 1,1-Dichloroethene-d2	2.063	65	22794	4.311	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.200%
20) 2-Butanone-d5	3.844	46	72225	40.766	ug/L	0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.540%
24) Chloroform-d	4.262	84	96031	4.836	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.800%
26) 1,2-Dichloroethane-d4	4.957	65	47650	4.720	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.400%
32) Benzene-d6	4.969	84	158408	4.558	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.200%
36) 1,2-Dichloropropane-d6	5.998	67	46103	4.562	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.200%
41) Toluene-d8	7.252	98	130844	3.954	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.000%
43) trans-1,3-Dichloroprop...	7.564	79	17108	4.516	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.400%
46) 2-Hexanone-d5	8.037	63	60780	45.795	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.600%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	35129	4.796	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.000%
66) 1,2-Dichlorobenzene-d4	11.567	152	56333	4.965	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.200%
Target Compounds						Qvalue
16) Methylene chloride	2.458	84	2309	0.250	ug/L	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050323\
Data File : VV030981.D
Acq On : 03 May 2023 10:53
Operator : SY/MD
Sample : VV0503WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK332

Quant Time: May 04 05:37:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu May 04 05:37:14 2023
Response via : Initial Calibration

