




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050422\
Data File : VV025840.D
Acq On : 04 May 2022 11:35
Operator : SY/MD
Sample : N2678-10
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH0E1

Manual IntegrationsAPPROVED

Reviewed By :John Carlone 05/05/2022
Supervised By :Semsettin Yesilyurt 05/06/2022

Quant Time: May 05 04:56:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR050322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu May 05 04:55:53 2022
Response via : Initial Calibration



TIC: VV025840.D\data.ms

(21) 2-Butanone (T)**3.998min (+ 0.026) 3.62 ug/L****response 7952**

Ion	Exp%	Act%
43.05	100.00	100.00
72.10	30.40	25.57
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050422\
Data File : VV025840.D
Acq On : 04 May 2022 11:35
Operator : SY/MD
Sample : N2678-10
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH0E1

Manual IntegrationsAPPROVED

Reviewed By :John Carlone 05/05/2022
Supervised By :Semsettin Yesilyurt 05/06/2022

Quant Time: May 05 04:56:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR050322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu May 05 04:55:53 2022
Response via : Initial Calibration



TIC: VV025840.D\data.ms

(21) 2-Butanone (T)

4.011min (+ 0.039) 7.97 ug/L m

response 17510

Ion	Exp%	Act%
43.05	100.00	100.00
72.10	30.40	11.61#
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050422\
 Data File : VV025840.D
 Acq On : 04 May 2022 11:35
 Operator : SY/MD
 Sample : N2678-10
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH0E1

Manual IntegrationsAPPROVED

Quant Time: May 05 04:56:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR050322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu May 05 04:55:53 2022
 Response via : Initial Calibration

Reviewed By :John Carlone 05/05/2022
 Supervised By :Semsettin Yesilyurt 05/06/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	166217	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	168369	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	79375	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	68855	3.779	ug/L	0.00
Spiked Amount 5.000	Range 40	- 130	Recovery	=	75.600%	
7) Chloroethane-d5	1.568	69	66868	4.173	ug/L	0.00
Spiked Amount 5.000	Range 65	- 130	Recovery	=	83.400%	
11) 1,1-Dichloroethene-d2	2.108	63	104567	3.021	ug/L	0.00
Spiked Amount 5.000	Range 60	- 125	Recovery	=	60.400%	
20) 2-Butanone-d5	3.896	46	104502	41.913	ug/L	0.00
Spiked Amount 50.000	Range 40	- 130	Recovery	=	83.820%	
24) Chloroform-d	4.346	84	113017	4.141	ug/L	0.00
Spiked Amount 5.000	Range 70	- 125	Recovery	=	82.800%	
26) 1,2-Dichloroethane-d4	5.031	65	52172	4.320	ug/L	0.00
Spiked Amount 5.000	Range 70	- 130	Recovery	=	86.400%	
32) Benzene-d6	5.047	84	226456	4.188	ug/L	0.00
Spiked Amount 5.000	Range 70	- 125	Recovery	=	83.800%	
36) 1,2-Dichloropropane-d6	6.066	67	68217	4.290	ug/L	0.00
Spiked Amount 5.000	Range 60	- 140	Recovery	=	85.800%	
41) Toluene-d8	7.313	98	181112	3.703	ug/L	0.00
Spiked Amount 5.000	Range 70	- 130	Recovery	=	74.000%	
43) trans-1,3-Dichloroprop...	7.622	79	18583	3.691	ug/L	0.00
Spiked Amount 5.000	Range 55	- 130	Recovery	=	73.800%	
46) 2-Hexanone-d5	8.088	63	80073	40.814	ug/L	0.00
Spiked Amount 50.000	Range 45	- 130	Recovery	=	81.620%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	42218	4.205	ug/L	0.00
Spiked Amount 5.000	Range 65	- 120	Recovery	=	84.000%	
66) 1,2-Dichlorobenzene-d4	11.622	152	63803	4.583	ug/L	0.00
Spiked Amount 5.000	Range 80	- 120	Recovery	=	91.600%	
Target Compounds						
13) Acetone	2.188	43	18474	11.258	ug/L	60
16) Methylene chloride	2.507	84	4578	0.295	ug/L	95
21) 2-Butanone	4.011	43	17510m	7.975	ug/L	
33) Benzene	5.101	78	10739	0.190	ug/L	100
42) Toluene	7.391	91	13805	0.230	ug/L	93
54) o-Xylene	9.548	106	1747	0.076	ug/L	73
63) 1,2,4-Trimethylbenzene	10.918	105	2572	0.061	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050422\
Data File : VV025840.D
Acq On : 04 May 2022 11:35
Operator : SY/MD
Sample : N2678-10
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH0E1

Manual IntegrationsAPPROVED

Quant Time: May 05 04:56:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR050322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu May 05 04:55:53 2022
Response via : Initial Calibration

Reviewed By :John Carlone 05/05/2022
Supervised By :Semsettin Yesilyurt 05/06/2022

