

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050423\
 Data File : VV030985.D
 Acq On : 04 May 2023 11:30
 Operator : SY/MD
 Sample : VSTDCCC0.5
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5296

Quant Time: May 05 05:13:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM050223.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 03 06:25:35 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	3807	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	3582	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2060	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	1654	0.512	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 102.000%	
4) 1,2-Dichloroethane-d4	4.945	65	1406	0.511	ug/L	-0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	= 102.000%	
7) 1,2-Dichloropropane-d6	5.995	67	1254	0.528	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 106.000%	
8) Toluene-d8	7.253	98	4560	0.525	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 106.000%	
10) 1,1,2,2-Tetrachloroeth...	10.146	84	955	0.513	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 102.000%	
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.280	62	1669	0.567	ug/L	99
6) Trichloroethene	5.852	95	1616	0.573	ug/L	98
9) 1,2-Dibromoethane	8.290	107	954	0.591	ug/L	98
12) 1,2,3-Trichloropropane	10.214	75	750	0.592	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.356	75	169	0.575	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050423\
Data File : VV030985.D
Acq On : 04 May 2023 11:30
Operator : SY/MD
Sample : VSTDCCC0.5
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5296

Quant Time: May 05 05:13:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM050223.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 03 06:25:35 2023
Response via : Initial Calibration

