

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050423\
 Data File : VV030987.D
 Acq On : 04 May 2023 12:52
 Operator : SY/MD
 Sample : 02215-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MDL-WATER-QT2-2023-04

Quant Time: May 05 05:15:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM050223.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 03 06:25:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	4026	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	3782	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2206	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	1537	0.450	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	90.000%		
4) 1,2-Dichloroethane-d4	4.970	65	1350	0.464	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	92.000%		
7) 1,2-Dichloropropane-d6	5.996	67	1246	0.497	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	100.000%		
8) Toluene-d8	7.254	98	4407	0.481	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	96.000%		
10) 1,1,2,2-Tetrachloroeth...	10.146	84	974	0.496	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	100.000%		
Target Compounds						Qvalue
3) Vinyl chloride	1.280	62	65	0.021	ug/L #	13
6) Trichloroethene	5.852	95	106	0.036	ug/L	90
9) 1,2-Dibromoethane	8.305	107	35	0.021	ug/L #	75
12) 1,2,3-Trichloropropane	10.214	75	33	0.024	ug/L #	45
13) 1,2-Dibromo-3-chloropr...	12.357	75	10	0.032	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050423\
Data File : VV030987.D
Acq On : 04 May 2023 12:52
Operator : SY/MD
Sample : 02215-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MDL-WATER-QT2-2023-04

Quant Time: May 05 05:15:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM050223.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 03 06:25:35 2023
Response via : Initial Calibration

