

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW050523\
 Data File : VW030995.D
 Acq On : 05 May 2023 15:09
 Operator : SY/MD
 Sample : 02613-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AA2

Quant Time: May 06 01:47:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat May 06 01:44:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	128508	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	128878	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	64701	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	39905	3.769	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.400%
7) Chloroethane-d5	1.539	69	35788	4.144	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.800%
11) 1,1-Dichloroethene-d2	2.063	65	21446	3.827	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.600%
20) 2-Butanone-d5	3.825	46	98918	52.685	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.380%
24) Chloroform-d	4.262	84	92610	4.401	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.000%
26) 1,2-Dichloroethane-d4	4.953	65	48983	4.579	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.600%
32) Benzene-d6	4.973	84	149079	3.946	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.000%
36) 1,2-Dichloropropane-d6	5.998	67	45829	4.172	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.400%
41) Toluene-d8	7.252	98	124451	3.459	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	69.200%#
43) trans-1,3-Dichloroprop...	7.564	79	16960	4.119	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.400%
46) 2-Hexanone-d5	8.037	63	69332	48.058	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.120%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	38300	4.810	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.200%
66) 1,2-Dichlorobenzene-d4	11.567	152	55938	4.638	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050523\
Data File : VV030995.D
Acq On : 05 May 2023 15:09
Operator : SY/MD
Sample : 02613-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AA2

Quant Time: May 06 01:47:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat May 06 01:44:36 2023
Response via : Initial Calibration

