

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050523\
 Data File : VV030992.D
 Acq On : 05 May 2023 12:24
 Operator : SY/MD
 Sample : VV0505WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK333

Quant Time: May 06 01:46:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat May 06 01:44:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	135716	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	131834	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	66889	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	43557	3.895	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	78.000%	
7) Chloroethane-d5	1.536	69	38514	4.222	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	84.400%	
11) 1,1-Dichloroethene-d2	2.063	65	23611	3.990	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	79.800%	
20) 2-Butanone-d5	3.844	46	81959	41.334	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	82.660%	
24) Chloroform-d	4.259	84	97539	4.389	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	87.800%	
26) 1,2-Dichloroethane-d4	4.954	65	51045	4.518	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.400%	
32) Benzene-d6	4.970	84	161125	4.170	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	83.400%	
36) 1,2-Dichloropropane-d6	5.998	67	47767	4.251	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	85.000%	
41) Toluene-d8	7.252	98	133797	3.636	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	72.800%	
43) trans-1,3-Dichloroprop...	7.564	79	17387	4.128	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	82.600%	
46) 2-Hexanone-d5	8.037	63	64468	43.685	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	87.360%	
56) 1,1,2,2-Tetrachloroeth...	10.162	84	36632	4.498	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	90.000%	
66) 1,2-Dichlorobenzene-d4	11.567	152	57501	4.612	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	92.200%	
Target Compounds						Qvalue
16) Methylene chloride	2.458	84	3178	0.308	ug/L	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050523\
Data File : VV030992.D
Acq On : 05 May 2023 12:24
Operator : SY/MD
Sample : VV0505WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK333

Quant Time: May 06 01:46:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat May 06 01:44:36 2023
Response via : Initial Calibration





#16
Methylene chloride
Concen: 0.308 ug/L
RT: 2.458 min Scan# 441
Delta R.T. 0.007 min
Lab File: VV030992.D
Acq: 05 May 2023 12:24

Instrument :
MSVOA_V
ClientSampleId :
VBLK333

Tgt Ion:	84	Resp:	3178
Ion	Ratio	Lower	Upper
84	100		
86	66.3	46.1	85.5
49	123.8	94.3	175.1

