

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050619\
 Data File : VV010675.D
 Acq On : 06 May 2019 13:51
 Operator : SY/MD
 Sample : K2603-22
 Misc : 5.06G/10ML/MSVOA V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GAJ74

Quant Time: May 07 04:17:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM050219S.M
 Quant Title : VOC Analysis
 QLast Update : Tue May 07 04:01:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	219536	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	201559	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	88276	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	39090	14.06	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	56.24%
7) Chloroethane-d5	1.58	69	31761	16.32	ug/L	0.02
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.28%
10) 1,1-Dichloroethene-d2	2.13	63	60189	12.35	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	49.40%
20) 2-Butanone-d5	3.93	46	19570	23.14	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	46.28%
24) Chloroform-d	4.40	84	102160	18.02	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	72.08%
26) 1,2-Dichloroethane-d4	5.08	65	53576	16.49	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	65.96%#
29) Benzene-d6	5.10	84	232297	22.24	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	88.96%
33) 1,2-Dichloropropane-d6	6.12	67	69944	21.86	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	87.44%
37) Toluene-d8	7.36	98	224719	22.39	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	89.56%
38) trans-1,3-Dichloropropene-	7.66	79	23305	17.62	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	70.48%
39) 2-Hexanone-d5	8.13	63	17510	30.74	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	61.48%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	45182	15.67	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	62.68%
61) 1,2-Dichlorobenzene-d4	11.67	152	74864	22.04	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	88.16%

Target Compounds

					Ovalue
13) Acetone	2.19	43	1735	2.393 ug/L	96
54) m,p-Xylene	9.18	106	3265	0.716 ug/L	96
55) o-xylene	9.59	106	3068	0.689 ug/L	85
56) Styrene	9.61	104	4378	0.579 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV050619\
Data File : VV010675.D
Acq On : 06 May 2019 13:51
Operator : SY/MD
Sample : K2603-22
Misc : 5.06G/10ML/MSVOA V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GAJ74

Quant Time: May 07 04:17:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM050219S.M
Quant Title : VOC Analysis
QLast Update : Tue May 07 04:01:38 2019
Response via : Initial Calibration

