

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050620\
 Data File : VV015916.D
 Acq On : 06 May 2020 17:40
 Operator : SY/MD
 Sample : L2527-05
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFSJ2

Quant Time: May 07 04:31:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050520WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 07 02:01:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	150028	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	148068	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	68595	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	39419	4.27	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.40%
7) Chloroethane-d5	1.58	69	37323	4.55	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.00%
11) 1,1-Dichloroethene-d2	2.12	63	60750	3.27	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.40%
20) 2-Butanone-d5	3.93	46	113779	47.56	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.12%
24) Chloroform-d	4.38	84	81751	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.40%
26) 1,2-Dichloroethane-d4	5.06	65	49423	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.20%
32) Benzene-d6	5.08	84	168656	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.00%
36) 1,2-Dichloropropane-d6	6.10	67	52327	4.62	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.40%
41) Toluene-d8	7.34	98	147616	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.40%
43) trans-1,3-Dichloropropene-	7.64	79	17949	4.04	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.80%
46) 2-Hexanone-d5	8.11	63	85788	43.58	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.16%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	36681	4.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.80%
64) 1,2-Dichlorobenzene-d4	11.65	152	55202	4.83	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.60%

Target Compounds

					Ovalue
13) Acetone	2.20	43	6287	4.268 ug/L	95
25) Chloroform	4.41	83	8274	0.421 ug/L	100
34) Trichloroethene	5.94	95	16455	1.559 ug/L	96
47) Tetrachloroethene	8.00	164	4965	0.631 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV050620\
Data File : VV015916.D
Acq On : 06 May 2020 17:40
Operator : SY/MD
Sample : L2527-05
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFSJ2

Quant Time: May 07 04:31:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050520WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 07 02:01:25 2020
Response via : Initial Calibration

