

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050620\
 Data File : VV015920.D
 Acq On : 06 May 2020 19:14
 Operator : SY/MD
 Sample : L2527-09
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFSJ6

Quant Time: May 07 04:45:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050520WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 07 02:01:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	151281	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	146375	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	68511	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	42565	4.58	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.60%
7) Chloroethane-d5	1.58	69	38477	4.65	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.00%
11) 1,1-Dichloroethene-d2	2.12	63	65124	3.48	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.60%
20) 2-Butanone-d5	3.93	46	118250	49.02	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.04%
24) Chloroform-d	4.38	84	84295	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.40%
26) 1,2-Dichloroethane-d4	5.06	65	52020	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.40%
32) Benzene-d6	5.08	84	176381	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.20%
36) 1,2-Dichloropropane-d6	6.10	67	54998	4.92	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.40%
41) Toluene-d8	7.34	98	158814	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.00%
43) trans-1,3-Dichloropropene-	7.64	79	18234	4.15	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.00%
46) 2-Hexanone-d5	8.11	63	88402	45.43	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.86%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	36898	4.56	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.20%
64) 1,2-Dichlorobenzene-d4	11.65	152	58459	5.12	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.40%

Target Compounds

					Ovalue
13) Acetone	2.20	43	7882	5.307	ug/L 69
25) Chloroform	4.41	83	12033	0.608	ug/L 97
30) Cyclohexane	4.71	56	4287	0.251	ug/L 96
34) Trichloroethene	5.95	95	4547	0.436	ug/L 95
47) Tetrachloroethene	8.00	164	33716	4.332	ug/L 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV050620\
Data File : VV015920.D
Acq On : 06 May 2020 19:14
Operator : SY/MD
Sample : L2527-09
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFSJ6

Quant Time: May 07 04:45:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050520WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 07 02:01:25 2020
Response via : Initial Calibration

