

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050719\
 Data File : VV010720.D
 Acq On : 08 May 2019 09:39
 Operator : SY/MD
 Sample : K2690-22MS
 Misc : 4.99G/10ML/MSVOA V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A5167MS

Quant Time: May 09 02:39:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM050719S.M
 Quant Title : VOC Analysis
 QLast Update : Thu May 09 02:24:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	110220	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	107663	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	41755	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	29894	17.46	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	69.84%
7) Chloroethane-d5	1.56	69	29016	20.59	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.36%
10) 1,1-Dichloroethene-d2	2.12	63	72805	19.60	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	78.40%
20) 2-Butanone-d5	3.92	46	27964	54.43	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	108.86%
24) Chloroform-d	4.39	84	72750	24.21	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	96.84%
26) 1,2-Dichloroethane-d4	5.07	65	48308	25.74	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	102.96%
29) Benzene-d6	5.09	84	135551	23.92	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.68%
33) 1,2-Dichloropropane-d6	6.11	67	45888	25.41	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.64%
37) Toluene-d8	7.36	98	117291	21.61	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	86.44%
38) trans-1,3-Dichloropropene-	7.66	79	16896	20.53	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	82.12%
39) 2-Hexanone-d5	8.13	63	21311	56.40	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	112.80%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	44867	26.44	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	105.76%
61) 1,2-Dichlorobenzene-d4	11.67	152	39827	24.63	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.52%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.13	96	39540	22.787	ug/L 94
13) Acetone	2.18	43	5106	6.419	ug/L 93
16) Methylene chloride	2.52	84	3232	1.896	ug/L 84
34) Benzene	5.14	78	155704	24.636	ug/L 100
35) Trichloroethene	5.96	95	39602	22.369	ug/L 98
44) Toluene	7.43	91	162457	22.632	ug/L 98
52) Chlorobenzene	8.92	112	104720	22.330	ug/L 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV050719\
Data File : VV010720.D
Acq On : 08 May 2019 09:39
Operator : SY/MD
Sample : K2690-22MS
Misc : 4.99G/10ML/MSVOA V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A5167MS

Quant Time: May 09 02:39:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM050719S.M
Quant Title : VOC Analysis
QLast Update : Thu May 09 02:24:26 2019
Response via : Initial Calibration

