

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050719\
 Data File : VV010727.D
 Acq On : 08 May 2019 14:29
 Operator : SY/MD
 Sample : VV0507SBL03
 Misc : 5.00G/10ML/MSVOA V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK83

Quant Time: May 09 03:02:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM050719S.M
 Quant Title : VOC Analysis
 QLast Update : Thu May 09 02:55:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	96641	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	94981	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	42555	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	42485	28.30	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	113.20%
7) Chloroethane-d5	1.56	69	33926	27.46	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	109.84%
10) 1,1-Dichloroethene-d2	2.12	63	67236	20.65	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	82.60%
20) 2-Butanone-d5	3.93	46	20827	46.23	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.46%
24) Chloroform-d	4.40	84	58744	22.30	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.20%
26) 1,2-Dichloroethane-d4	5.08	65	40548	24.64	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	98.56%
29) Benzene-d6	5.09	84	134207	26.85	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	107.40%
33) 1,2-Dichloropropane-d6	6.12	67	40893	25.67	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	102.68%
37) Toluene-d8	7.36	98	125585	26.23	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	104.92%
38) trans-1,3-Dichloropropene-	7.66	79	17013	23.43	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	93.72%
39) 2-Hexanone-d5	8.13	63	15757	47.27	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.54%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	35164	23.49	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	93.96%
61) 1,2-Dichlorobenzene-d4	11.67	152	41592	25.24	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	100.96%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV050719\
Data File : VV010727.D
Acq On : 08 May 2019 14:29
Operator : SY/MD
Sample : VV0507SBL03
Misc : 5.00G/10ML/MSVOA V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK83

Quant Time: May 09 03:02:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM050719S.M
Quant Title : VOC Analysis
QLast Update : Thu May 09 02:55:09 2019
Response via : Initial Calibration

