

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050719\
 Data File : VV010742.D
 Acq On : 08 May 2019 23:03
 Operator : SY/MD
 Sample : K2690-20
 Misc : 4.96G/10ML/MSVOA V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A5166

Quant Time: May 09 04:42:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM050719S.M
 Quant Title : VOC Analysis
 QLast Update : Thu May 09 02:55:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	81629	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	84874	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	31909	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	22474	17.72	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	70.88%
7) Chloroethane-d5	1.57	69	24216	23.21	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.84%
10) 1,1-Dichloroethene-d2	2.13	63	39008	14.18	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	56.72%
20) 2-Butanone-d5	3.93	46	23991	63.05	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	126.10%
24) Chloroform-d	4.40	84	57832	25.99	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	103.96%
26) 1,2-Dichloroethane-d4	5.08	65	40517	29.15	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	116.60%
29) Benzene-d6	5.10	84	105690	23.66	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	94.64%
33) 1,2-Dichloropropane-d6	6.12	67	37614	26.42	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	105.68%
37) Toluene-d8	7.36	98	88534	20.69	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	82.76%
38) trans-1,3-Dichloropropene-	7.66	79	12366	19.06	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	76.24%
39) 2-Hexanone-d5	8.13	63	17736	59.54	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	119.08%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	39626	29.62	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	118.48%
61) 1,2-Dichlorobenzene-d4	11.67	152	29886	24.18	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.72%

Target Compounds					Ovalue
13) Acetone	2.19	43	5183	8.798 ug/L	100
15) Methyl Acetate	2.47	43	1156	1.643 ug/L	93
16) Methylene chloride	2.54	84	2674	2.119 ug/L	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV050719\
Data File : VV010742.D
Acq On : 08 May 2019 23:03
Operator : SY/MD
Sample : K2690-20
Misc : 4.96G/10ML/MSVOA V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A5166

Quant Time: May 09 04:42:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM050719S.M
Quant Title : VOC Analysis
QLast Update : Thu May 09 02:55:09 2019
Response via : Initial Calibration

