

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050719\
 Data File : VV010695.D
 Acq On : 07 May 2019 16:48
 Operator : SY/MD
 Sample : VV0507SBL01
 Misc : 5.00G/10ML/MSVOA V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK81

Quant Time: May 08 03:41:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM050719S.M
 Quant Title : VOC Analysis
 QLast Update : Wed May 08 03:02:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	94856	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	97738	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	44368	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	36443	24.73	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	98.92%
7) Chloroethane-d5	1.56	69	30676	25.30	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.20%
10) 1,1-Dichloroethene-d2	2.12	63	57080	17.86	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	71.44%
20) 2-Butanone-d5	3.93	46	23191	52.45	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	104.90%
24) Chloroform-d	4.40	84	60722	23.48	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.92%
26) 1,2-Dichloroethane-d4	5.08	65	41511	25.70	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	102.80%
29) Benzene-d6	5.09	84	123104	23.93	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.72%
33) 1,2-Dichloropropane-d6	6.12	67	40426	24.66	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	98.64%
37) Toluene-d8	7.36	98	117402	23.83	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	95.32%
38) trans-1,3-Dichloropropene-	7.66	79	17727	23.72	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	94.88%
39) 2-Hexanone-d5	8.13	63	18714	54.56	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	109.12%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	37080	24.07	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	96.28%
61) 1,2-Dichlorobenzene-d4	11.67	152	40953	23.83	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.32%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV050719\
Data File : VV010695.D
Acq On : 07 May 2019 16:48
Operator : SY/MD
Sample : VV0507SBL01
Misc : 5.00G/10ML/MSVOA V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK81

Quant Time: May 08 03:41:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM050719S.M
Quant Title : VOC Analysis
QLast Update : Wed May 08 03:02:12 2019
Response via : Initial Calibration

