

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050721\
 Data File : VV021332.D
 Acq On : 07 May 2021 17:46
 Operator : SY/MD
 Sample : M2320-02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: May 08 00:38:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM050321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat May 08 00:36:06 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	680289	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	667066	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.255	152	356765	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	138935	46.04	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	92.08%	
7) Chloroethane-d5	1.571	69	127847	47.41	ug/L	0.00
Spiked Amount 50.000	Range 70	- 130	Recovery	=	94.82%	
11) 1,1-Dichloroethene-d2	2.108	63	275608	32.74	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	65.48%	
21) 2-Butanone-d5	3.896	46	208111	85.94	ug/L	0.00
Spiked Amount 100.000	Range 40	- 130	Recovery	=	85.94%	
24) Chloroform-d	4.349	84	418901	43.61	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	87.22%	
26) 1,2-Dichloroethane-d4	5.037	65	335059	46.60	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	93.20%	
32) Benzene-d6	5.056	84	730411	47.54	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	95.08%	
36) 1,2-Dichloropropane-d6	6.076	67	185043m	46.89	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	93.78%	
41) Toluene-d8	7.320	98	743189	46.68	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	93.36%	
43) trans-1,3-Dichloroprop...	7.625	79	127931	43.58	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	87.16%	
47) 2-Hexanone-d5	8.092	63	158388	86.91	ug/L	0.00
Spiked Amount 100.000	Range 45	- 130	Recovery	=	86.91%	
56) 1,1,2,2-Tetrachloroeth...	10.220	84	290298	45.97	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	91.94%	
66) 1,2-Dichlorobenzene-d4	11.632	152	335291	47.69	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	95.38%	
Target Compounds						
17) trans-1,2-Dichloroethene	2.767	96	44276	10.38	ug/L	92
18) Methyl tert-butyl Ether	2.770	73	4313302	282.07	ug/L #	80
20) cis-1,2-Dichloroethene	3.918	96	32611	6.95	ug/L	99
29) Cyclohexane	4.680	56	23655	4.22	ug/L #	64
33) Benzene	5.108	78	27044	1.58	ug/L	100
34) Trichloroethene	5.924	95	15780	2.87	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed