

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050820\
 Data File : VV015959.D
 Acq On : 08 May 2020 12:50
 Operator : SY/MD
 Sample : L2520-11
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H41W9

Quant Time: May 09 01:49:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat May 09 00:44:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	141129	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	140532	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	65300	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	36090	4.71	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.20%
7) Chloroethane-d5	1.58	69	37549	5.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.40%
11) 1,1-Dichloroethene-d2	2.12	63	60461	3.67	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.40%
20) 2-Butanone-d5	3.92	46	120635	55.24	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.48%
24) Chloroform-d	4.38	84	91238	5.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.00%
26) 1,2-Dichloroethane-d4	5.06	65	52088	5.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.60%
32) Benzene-d6	5.08	84	174400	5.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.40%
36) 1,2-Dichloropropane-d6	6.10	67	55429	5.43	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.60%
41) Toluene-d8	7.34	98	157766	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.80%
45) trans-1,3-Dichloropropene-	7.65	79	18926	4.73	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.60%
48) 2-Hexanone-d5	8.12	63	88467	50.27	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.54%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	37534	5.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.60%
65) 1,2-Dichlorobenzene-d4	11.65	152	59834	5.65	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.00%

Target Compounds					Ovalue
13) Acetone	2.19	43	4681	3.249 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050820\
Data File : VV015959.D
Acq On : 08 May 2020 12:50
Operator : SY/MD
Sample : L2520-11
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H41W9

Quant Time: May 09 01:49:00 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat May 09 00:44:25 2020
Response via : Initial Calibration

