

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050820\
 Data File : VV015975.D
 Acq On : 08 May 2020 19:03
 Operator : SY/MD
 Sample : L2520-16
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H4222

Quant Time: May 09 04:48:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat May 09 00:44:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	142263	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	141136	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	65927	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	32436	4.20	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.00%
7) Chloroethane-d5	1.58	69	33446	4.57	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.40%
11) 1,1-Dichloroethene-d2	2.12	63	54056	3.26	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.20%
20) 2-Butanone-d5	3.92	46	109271	49.63	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.26%
24) Chloroform-d	4.38	84	79013	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.60%
26) 1,2-Dichloroethane-d4	5.06	65	47656	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.60%
32) Benzene-d6	5.08	84	158951	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.60%
36) 1,2-Dichloropropane-d6	6.10	67	50054	4.88	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.60%
41) Toluene-d8	7.34	98	140289	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.40%
45) trans-1,3-Dichloropropene-	7.65	79	15779	3.92	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.40%
48) 2-Hexanone-d5	8.11	63	81994	46.39	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.78%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	35844	4.97	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.40%
65) 1,2-Dichlorobenzene-d4	11.65	152	54187	5.06	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.20%

Target Compounds

					Ovalue
13) Acetone	2.20	43	2961	2.039	ug/L 68
22) cis-1,2-Dichloroethene	3.95	96	9974	1.101	ug/L 93
25) Chloroform	4.41	83	6755	0.368	ug/L 98
34) Trichloroethene	5.94	95	33707	3.326	ug/L 96
49) Tetrachloroethene	8.00	164	22403	3.168	ug/L 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV050820\
Data File : VV015975.D
Acq On : 08 May 2020 19:03
Operator : SY/MD
Sample : L2520-16
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H4222

Quant Time: May 09 04:48:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat May 09 00:44:25 2020
Response via : Initial Calibration

