


Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV050820\
 Data File : VV015951.D
 Acq On : 08 May 2020 08:58
 Operator : SY/MD
 Sample : BFB53
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB53

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Sat May 09 00:44:25 2020



AutoFind: Scans 2800, 2801, 2802; Background Corrected with Scan 2788

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.3	25442	PASS
75	95	30	80	54.0	61674	PASS
95	95	100	100	100.0	114290	PASS
96	95	5	9	6.8	7782	PASS
173	174	0.00	2	1.4	1127	PASS
174	95	50	120	70.7	80802	PASS
175	174	5	9	7.6	6110	PASS
176	174	95	101	96.8	78205	PASS
177	176	5	9	6.7	5202	PASS

Scan 2788 (10.055 min): VV015951.D
BFB53

m/z	Abundance
40.00	363.0
43.90	407.0

Instrument :
MSVOA_V
ClientSampleId :
BFB53

m/z	Abundance
36.00	1239.0
37.00	6757.0
38.00	6103.0
39.00	2407.0
39.90	400.0
43.00	145.0
44.00	1119.0
45.00	1309.0
47.00	1666.0
47.90	824.0
49.00	5439.0
50.00	25128.0
51.00	7200.0
52.00	325.0
52.90	178.0
54.90	426.0
56.00	1996.0
57.00	3629.0
58.00	174.0
60.00	1313.0
61.00	6056.0
62.00	6098.0
63.00	4882.0
64.00	528.0
65.00	796.0
65.90	138.0
67.00	363.0
68.00	12485.0
69.00	12878.0
70.00	915.0
72.00	687.0
73.00	5932.0
74.00	22376.0
75.00	62584.0
76.00	5114.0
77.00	710.0
77.90	481.0
78.90	4362.0
79.90	1376.0
80.90	4620.0
81.90	997.0
82.90	163.0
86.10	126.0
87.00	4615.0
87.90	4208.0
90.80	475.0
92.00	3620.0
93.00	5245.0
94.00	14946.0
95.00	115472.0
96.00	7411.0
97.00	233.0
103.90	585.0
104.90	185.0
105.90	655.0
106.90	110.0
110.70	177.0
111.70	108.0
112.80	122.0
114.80	211.0
115.80	615.0
116.80	913.0
117.80	508.0
118.80	725.0
127.90	490.0
128.90	249.0
129.90	428.0
130.80	183.0
134.90	294.0
136.90	224.0
140.90	1109.0
141.80	190.0
142.90	1457.0
145.00	197.0
145.80	205.0
147.80	338.0
149.70	122.0
152.90	132.0
153.90	116.0
154.80	336.0
156.90	207.0
158.80	139.0
160.90	111.0
170.80	182.0
171.90	873.0
173.00	1275.0
173.90	83680.0
175.00	6298.0
175.90	82088.0
176.90	5339.0
178.00	164.0

Instrument :
MSVOA_V
ClientSampleId :
BFB53

m/z	Abundance
36.00	1291.0
37.00	6995.0
38.00	6377.0
39.00	2495.0
39.90	384.0
40.90	149.0
41.80	110.0
43.00	153.0
44.00	1128.0
45.00	1285.0
46.10	133.0
47.00	1718.0
47.90	903.0
49.00	5543.0
50.00	26472.0
51.00	7213.0
52.00	352.0
52.90	166.0
55.00	472.0
56.00	2049.0
57.00	3658.0
58.00	199.0
60.00	1350.0
61.00	6331.0
62.00	6398.0
63.00	4990.0
64.00	515.0
65.00	872.0
65.90	113.0
67.00	356.0
68.00	13205.0
69.00	13258.0
70.00	1076.0
72.00	695.0
73.00	6068.0
74.00	22368.0
75.00	63720.0
76.00	5370.0
77.00	746.0
77.90	471.0
78.90	4428.0
79.90	1374.0
80.90	4744.0
81.90	1083.0
83.00	157.0
85.80	122.0
86.00	107.0
86.90	4469.0
87.90	4281.0
90.80	483.0
92.00	3616.0
93.00	5448.0
94.00	15333.0
95.00	118312.0
96.00	8079.0
96.90	259.0
103.90	589.0
104.80	170.0
106.00	656.0
106.90	111.0
110.70	148.0
111.70	117.0
112.90	137.0
114.90	199.0
115.80	604.0
116.90	935.0
117.90	513.0
118.80	755.0
127.90	475.0
128.90	201.0
129.90	464.0
130.80	183.0
135.00	258.0
136.80	244.0
139.80	117.0
140.90	1187.0
141.80	199.0
142.90	1448.0
144.70	210.0
145.80	205.0
147.80	304.0
149.70	114.0
153.90	117.0
154.90	340.0
156.90	183.0
158.90	128.0
170.90	181.0
171.90	870.0
173.00	1158.0
173.90	83512.0
175.00	6372.0
175.90	80840.0

Instrument :
MSVOA_V
ClientSampleId :
BFB53

176.905426.0

178.00154.0

Instrument :
MSVOA_V
ClientSampleId :
BFB53

m/z	Abundance
36.00	1223.0
37.00	6567.0
38.00	5937.0
39.00	2351.0
39.90	382.0
40.90	177.0
41.80	127.0
43.00	159.0
44.00	1061.0
45.00	1175.0
46.10	131.0
47.00	1598.0
47.90	896.0
49.00	5190.0
50.00	24728.0
51.00	6466.0
52.00	349.0
52.90	142.0
55.00	435.0
56.00	1844.0
57.00	3368.0
57.90	210.0
60.00	1258.0
61.00	6018.0
62.00	6072.0
63.00	4624.0
64.00	521.0
65.00	935.0
67.00	315.0
68.00	12292.0
69.00	12517.0
70.00	1097.0
72.00	668.0
73.00	5559.0
74.00	20464.0
75.00	58720.0
76.00	5157.0
76.90	618.0
77.90	438.0
78.90	4143.0
79.90	1282.0
80.90	4367.0
81.90	1117.0
82.90	149.0
85.80	126.0
86.90	4028.0
87.90	3876.0
90.80	449.0
92.00	3374.0
93.00	5052.0
94.00	13816.0
95.00	109088.0
96.00	7856.0
96.90	305.0
103.90	537.0
104.80	166.0
105.90	581.0
106.90	113.0
110.70	109.0
111.80	118.0
112.90	159.0
114.90	178.0
115.80	574.0
116.90	803.0
117.90	497.0
118.80	697.0
127.90	421.0
128.90	161.0
129.90	459.0
130.80	177.0
135.00	194.0
136.80	241.0
139.90	131.0
140.90	1212.0
141.80	182.0
142.90	1310.0
144.80	214.0
145.80	202.0
147.90	236.0
154.90	296.0
156.90	151.0
159.00	135.0
170.90	190.0
171.90	807.0
173.10	948.0
173.90	75216.0
174.90	5662.0
175.90	71688.0
176.90	4843.0
178.00	154.0

Instrument :
MSVOA_V
ClientSampleId :
BFB53