

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV051120\
 Data File : VV015981.D
 Acq On : 11 May 2020 11:54
 Operator : SY/MD
 Sample : VV0511WBL01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK59

Quant Time: May 12 01:20:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue May 12 01:07:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	141633	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	140513	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	63383	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	32294	4.20	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.00%
7) Chloroethane-d5	1.58	69	32711	4.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.80%
11) 1,1-Dichloroethene-d2	2.12	63	52598	3.18	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.60%
20) 2-Butanone-d5	3.96	46	116702	53.25	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.50%
24) Chloroform-d	4.38	84	82371	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
26) 1,2-Dichloroethane-d4	5.06	65	50880	5.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.60%
32) Benzene-d6	5.08	84	160578	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
36) 1,2-Dichloropropane-d6	6.09	67	51861	5.08	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.60%
41) Toluene-d8	7.34	98	142973	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.40%
45) trans-1,3-Dichloropropene-	7.65	79	19051	4.76	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.20%
48) 2-Hexanone-d5	8.12	63	86076	48.92	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.84%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	37666	5.25	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.00%
65) 1,2-Dichlorobenzene-d4	11.65	152	55003	5.35	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.00%

Target Compounds					Ovalue
25) Chloroform	4.41	83	5771	0.316 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV051120\
Data File : VV015981.D
Acq On : 11 May 2020 11:54
Operator : SY/MD
Sample : VV0511WBL01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK59

Quant Time: May 12 01:20:40 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue May 12 01:07:44 2020
Response via : Initial Calibration

