

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV051120\
 Data File : VV015985.D
 Acq On : 11 May 2020 13:28
 Operator : SY/MD
 Sample : L2520-08DL 5X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H41Y3DL

Quant Time: May 12 02:02:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue May 12 01:07:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	145798	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	140282	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	65388	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	29828	3.77	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.40%
7) Chloroethane-d5	1.58	69	30685	4.09	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.80%
11) 1,1-Dichloroethene-d2	2.12	63	51402	3.02	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.40%
20) 2-Butanone-d5	3.94	46	112639	49.92	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.84%
24) Chloroform-d	4.38	84	81645	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.40%
26) 1,2-Dichloroethane-d4	5.06	65	45788	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.40%
32) Benzene-d6	5.08	84	152013	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.00%
36) 1,2-Dichloropropane-d6	6.09	67	49635	4.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.40%
41) Toluene-d8	7.34	98	137952	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.40%
45) trans-1,3-Dichloropropene-	7.65	79	17580	4.40	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.00%
48) 2-Hexanone-d5	8.12	63	77559	44.15	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.30%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	34789	4.85	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.00%
65) 1,2-Dichlorobenzene-d4	11.65	152	53070	5.00	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.00%

Target Compounds					Ovalue
22) cis-1,2-Dichloroethene	3.94	96	22956	2.473 ug/L	97
25) Chloroform	4.41	83	15952	0.848 ug/L	96
34) Trichloroethene	5.94	95	4028	0.400 ug/L	93
49) Tetrachloroethene	8.00	164	45229	6.435 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV051120\
Data File : VV015985.D
Acq On : 11 May 2020 13:28
Operator : SY/MD
Sample : L2520-08DL 5X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H41Y3DL

Quant Time: May 12 02:02:02 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue May 12 01:07:44 2020
Response via : Initial Calibration

