

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051123\
 Data File : VV031012.D
 Acq On : 11 May 2023 10:46
 Operator : SY/MD
 Sample : 02697-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C11E8

Quant Time: May 12 09:46:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat May 06 01:44:36 2023
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		5.542	114	125194	5.000	ug/L	0.00
28) Chlorobenzene-d5		8.789	117	121609	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.194	152	60370	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.281	65	52698	5.109	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.200%	
7) Chloroethane-d5		1.539	69	48728	5.791	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	115.800%	
11) 1,1-Dichloroethene-d2		2.063	65	27639	5.063	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	101.200%	
20) 2-Butanone-d5		3.818	46	90568	49.515	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.020%	
24) Chloroform-d		4.259	84	114212	5.572	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.400%	
26) 1,2-Dichloroethane-d4		4.950	65	55944	5.368	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.400%	
32) Benzene-d6		4.966	84	200694	5.630	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.600%	
36) 1,2-Dichloropropane-d6		5.995	67	56588	5.459	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.200%	
41) Toluene-d8		7.249	98	172565	5.084	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.600%	
43) trans-1,3-Dichloroprop...		7.561	79	19963	5.138	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	102.800%	
46) 2-Hexanone-d5		8.037	63	64039	47.043	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.080%	
56) 1,1,2,2-Tetrachloroeth...		10.159	84	39083	5.202	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.000%	
66) 1,2-Dichlorobenzene-d4		11.567	152	63377	5.632	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.600%	
Target Compounds							
3) Chloromethane		1.217	50	9923	0.855	ug/L	99
13) Acetone		2.150	43	4138	2.648	ug/L	96
48) 2-Hexanone		8.095	43	9938	3.447	ug/L #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051123\
Data File : VV031012.D
Acq On : 11 May 2023 10:46
Operator : SY/MD
Sample : 02697-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C11E8

Quant Time: May 12 09:46:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat May 06 01:44:36 2023
Response via : Initial Calibration

