

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051222\
 Data File : VV025878.D
 Acq On : 12 May 2022 13:21
 Operator : SY/MD
 Sample : VV0512WBL01
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK214

Quant Time: May 13 07:08:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 13 07:07:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	317256	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	313215	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	144915	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	159736	39.012	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	78.020%
7) Chloroethane-d5	1.571	69	133392	37.098	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	74.200%
11) 1,1-Dichloroethene-d2	2.108	63	237134	33.114	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.220%
21) 2-Butanone-d5	3.905	46	150459	85.734	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.730%
24) Chloroform-d	4.349	84	244971	42.835	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.680%
26) 1,2-Dichloroethane-d4	5.031	65	157148	45.467	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.940%
32) Benzene-d6	5.050	84	487327	44.630	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.260%
36) 1,2-Dichloropropane-d6	6.069	67	157090	46.841	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.680%
41) Toluene-d8	7.313	98	417029	41.452	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.900%
43) trans-1,3-Dichloroprop...	7.622	79	64302	42.952	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.900%
47) 2-Hexanone-d5	8.088	63	111949	91.049	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	91.050%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	204029	45.620	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.240%
66) 1,2-Dichlorobenzene-d4	11.622	152	157141	49.434	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.860%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051222\
Data File : VV025878.D
Acq On : 12 May 2022 13:21
Operator : SY/MD
Sample : VV0512WBL01
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK214

Quant Time: May 13 07:08:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 13 07:07:32 2022
Response via : Initial Calibration

