

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW051222\  
 Data File : VW025882.D  
 Acq On : 12 May 2022 15:09  
 Operator : SY/MD  
 Sample : N2714-06ME  
 Misc : 3.55g/5.0mL/100uL/5.00mL/MSVOA\_V/MEOH  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 EW2X7ME

Quant Time: May 13 07:09:17 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVLM051022WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Fri May 13 07:07:32 2022  
 Response via : Initial Calibration

| Compound                      | R.T.    | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|---------|-------|----------|----------|-------|----------|
| -----                         |         |       |          |          |       |          |
| Internal Standards            |         |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 5.616   | 114   | 316078   | 50.000   | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 8.850   | 117   | 315279   | 50.000   | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.249  | 152   | 143756   | 50.000   | ug/L  | 0.00     |
| System Monitoring Compounds   |         |       |          |          |       |          |
| 4) Vinyl Chloride-d3          | 1.307   | 65    | 147467   | 36.150   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 60 - 135 | Recovery | =     | 72.300%  |
| 7) Chloroethane-d5            | 1.568   | 69    | 125285   | 34.973   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 130 | Recovery | =     | 69.940%# |
| 11) 1,1-Dichloroethene-d2     | 2.108   | 63    | 221667   | 31.070   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 60 - 125 | Recovery | =     | 62.140%  |
| 21) 2-Butanone-d5             | 3.908   | 46    | 169994   | 97.227   | ug/L  | 0.01     |
| Spiked Amount                 | 100.000 | Range | 40 - 130 | Recovery | =     | 97.230%  |
| 24) Chloroform-d              | 4.349   | 84    | 230616   | 40.475   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 125 | Recovery | =     | 80.960%  |
| 26) 1,2-Dichloroethane-d4     | 5.031   | 65    | 151201   | 43.909   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 125 | Recovery | =     | 87.820%  |
| 32) Benzene-d6                | 5.050   | 84    | 459369   | 41.795   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 125 | Recovery | =     | 83.580%  |
| 36) 1,2-Dichloropropane-d6    | 6.069   | 67    | 148855   | 44.095   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 120 | Recovery | =     | 88.180%  |
| 41) Toluene-d8                | 7.317   | 98    | 394749   | 38.980   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 80 - 120 | Recovery | =     | 77.960%# |
| 43) trans-1,3-Dichloroprop... | 7.622   | 79    | 61584    | 40.867   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 60 - 125 | Recovery | =     | 81.740%  |
| 47) 2-Hexanone-d5             | 8.092   | 63    | 113824   | 91.968   | ug/L  | 0.00     |
| Spiked Amount                 | 100.000 | Range | 45 - 130 | Recovery | =     | 91.970%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.214  | 84    | 197701   | 43.916   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 65 - 120 | Recovery | =     | 87.840%  |
| 66) 1,2-Dichlorobenzene-d4    | 11.622  | 152   | 148549   | 47.108   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 80 - 120 | Recovery | =     | 94.220%  |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV051222\  
Data File : VV025882.D  
Acq On : 12 May 2022 15:09  
Operator : SY/MD  
Sample : N2714-06ME  
Misc : 3.55g/5.0mL/100uL/5.00mL/MSVOA\_V/MEOH  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
EW2X7ME

Quant Time: May 13 07:09:17 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVLM051022WMA.M  
Quant Title : VOC Analysis  
QLast Update : Fri May 13 07:07:32 2022  
Response via : Initial Calibration

