

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW051222\
 Data File : VW025886.D
 Acq On : 12 May 2022 16:44
 Operator : SY/MD
 Sample : N2713-16
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBSL5

Quant Time: May 13 07:10:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 13 07:07:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	283323	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	289567	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	128725	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	145218	39.714	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	79.420%	
7) Chloroethane-d5	1.568	69	125273	39.012	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	78.020%	
11) 1,1-Dichloroethene-d2	2.108	63	222363	34.770	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	69.540%	
21) 2-Butanone-d5	3.908	46	166080	105.970	ug/L	0.01
Spiked Amount	100.000	Range 40 - 130	Recovery	=	105.970%	
24) Chloroform-d	4.349	84	227154	44.477	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	88.960%	
26) 1,2-Dichloroethane-d4	5.034	65	150347	48.709	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.420%	
32) Benzene-d6	5.050	84	458508	45.420	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	90.840%	
36) 1,2-Dichloropropane-d6	6.069	67	148416	47.869	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	95.740%	
41) Toluene-d8	7.317	98	390795	42.017	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	84.040%	
43) trans-1,3-Dichloroprop...	7.622	79	61315	44.302	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	88.600%	
47) 2-Hexanone-d5	8.091	63	113574	99.914	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	99.910%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	196358	47.490	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	94.980%	
66) 1,2-Dichlorobenzene-d4	11.622	152	144288	51.099	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	102.200%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051222\
Data File : VV025886.D
Acq On : 12 May 2022 16:44
Operator : SY/MD
Sample : N2713-16
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBSL5

Quant Time: May 13 07:10:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 13 07:07:32 2022
Response via : Initial Calibration

