

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV051320\
 Data File : VV016035.D
 Acq On : 13 May 2020 13:02
 Operator : SY/MD
 Sample : L2542-13
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H4216

Quant Time: May 14 02:59:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 14 02:25:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	145894	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	145011	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	64553	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	48427	6.12	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	122.40%
7) Chloroethane-d5	1.58	69	44355	5.91	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	118.20%
11) 1,1-Dichloroethene-d2	2.12	63	70388	4.14	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.80%
20) 2-Butanone-d5	3.93	46	116381	51.55	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.10%
24) Chloroform-d	4.38	84	93776	5.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.40%
26) 1,2-Dichloroethane-d4	5.06	65	51702	5.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.20%
32) Benzene-d6	5.08	84	179814	5.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.20%
36) 1,2-Dichloropropane-d6	6.10	67	56226	5.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.80%
41) Toluene-d8	7.34	98	161298	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.80%
45) trans-1,3-Dichloropropene-	7.65	79	16911	4.09	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.80%
48) 2-Hexanone-d5	8.12	63	90419	49.79	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.58%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	38983	5.26	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.20%
65) 1,2-Dichlorobenzene-d4	11.65	152	57235	5.46	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.20%

Target Compounds

					Ovalue
13) Acetone	2.21	43	5538	3.718 ug/L	64
25) Chloroform	4.41	83	5067	0.269 ug/L	99
34) Trichloroethene	5.95	95	2605	0.250 ug/L	93
49) Tetrachloroethene	8.00	164	61367	8.446 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV051320\
Data File : VV016035.D
Acq On : 13 May 2020 13:02
Operator : SY/MD
Sample : L2542-13
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H4216

Quant Time: May 14 02:59:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 14 02:25:38 2020
Response via : Initial Calibration

