

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV051320\
 Data File : VV016046.D
 Acq On : 13 May 2020 17:15
 Operator : SY/MD
 Sample : L2542-14DL 40X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H4217DL

Quant Time: May 14 05:06:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 14 02:25:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	150119	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	148014	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	68325	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	45287	5.56	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	111.20%
7) Chloroethane-d5	1.58	69	41694	5.40	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.00%
11) 1,1-Dichloroethene-d2	2.13	63	66216	3.78	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.60%
20) 2-Butanone-d5	3.95	46	113831	49.00	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.00%
24) Chloroform-d	4.38	84	91247	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.60%
26) 1,2-Dichloroethane-d4	5.06	65	51069	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.00%
32) Benzene-d6	5.08	84	178194	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.20%
36) 1,2-Dichloropropane-d6	6.10	67	54484	5.07	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.40%
41) Toluene-d8	7.34	98	156792	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.20%
45) trans-1,3-Dichloropropene-	7.65	79	18785	4.45	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.00%
48) 2-Hexanone-d5	8.12	63	89867	48.48	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.96%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	38229	5.06	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.20%
65) 1,2-Dichlorobenzene-d4	11.65	152	55899	5.04	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.80%

Target Compounds					Ovalue
16) Methylene chloride	2.53	84	3579	0.362 ug/L	93
17) Methyl tert-butyl Ether	2.79	73	156173	7.275 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV051320\
Data File : VV016046.D
Acq On : 13 May 2020 17:15
Operator : SY/MD
Sample : L2542-14DL 40X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H4217DL

Quant Time: May 14 05:06:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 14 02:25:38 2020
Response via : Initial Calibration

