

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV051320\
 Data File : VV016054.D
 Acq On : 13 May 2020 20:45
 Operator : SY/MD
 Sample : L2598-08
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H4232

Quant Time: May 14 05:59:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 14 02:25:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	150443	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	150723	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	68508	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	49879	6.11	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	122.20%
7) Chloroethane-d5	1.58	69	46743	6.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	120.80%
11) 1,1-Dichloroethene-d2	2.13	63	87104	4.96	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	99.20%
20) 2-Butanone-d5	3.94	46	123144	52.89	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.78%
24) Chloroform-d	4.38	84	97829	5.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.60%
26) 1,2-Dichloroethane-d4	5.06	65	54437	5.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.40%
32) Benzene-d6	5.08	84	190747	5.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.40%
36) 1,2-Dichloropropane-d6	6.10	67	60066	5.49	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.80%
41) Toluene-d8	7.34	98	170273	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.40%
45) trans-1,3-Dichloropropene-	7.64	79	18318	4.26	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.20%
48) 2-Hexanone-d5	8.12	63	92582	49.05	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.10%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	40799	5.30	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.00%
65) 1,2-Dichlorobenzene-d4	11.65	152	60073	5.40	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.00%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.14	96	25122	3.034 ug/L	# 75
13) Acetone	2.22	43	3903	2.541 ug/L	89
22) cis-1,2-Dichloroethene	3.95	96	1930	0.202 ug/L	86
25) Chloroform	4.41	83	6574	0.339 ug/L	96
29) 1,1,1-Trichloroethane	4.64	97	7879	0.474 ug/L	98
34) Trichloroethene	5.94	95	5882	0.543 ug/L	95
49) Tetrachloroethene	8.00	164	50370	6.670 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV051320\
Data File : VV016054.D
Acq On : 13 May 2020 20:45
Operator : SY/MD
Sample : L2598-08
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H4232

Quant Time: May 14 05:59:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 14 02:25:38 2020
Response via : Initial Calibration

