

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV051420\
 Data File : VV016073.D
 Acq On : 14 May 2020 17:44
 Operator : SY/MD
 Sample : L2598-16
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: May 15 08:23:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 15 06:27:04 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	151825	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	151328	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	69061	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	40912	4.97	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.40%
7) Chloroethane-d5	1.58	69	40475	5.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.60%
11) 1,1-Dichloroethene-d2	2.12	63	61486	3.47	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.40%
20) 2-Butanone-d5	3.94	46	124918	53.17	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.34%
24) Chloroform-d	4.38	84	85642	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.20%
26) 1,2-Dichloroethane-d4	5.06	65	49819	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.60%
32) Benzene-d6	5.08	84	168402	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.60%
36) 1,2-Dichloropropane-d6	6.10	67	53399	4.86	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.20%
41) Toluene-d8	7.34	98	146979	4.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.40%
45) trans-1,3-Dichloropropene-	7.65	79	17742	4.11	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.20%
48) 2-Hexanone-d5	8.12	63	96296	50.82	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.64%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	38863	5.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.60%
65) 1,2-Dichlorobenzene-d4	11.65	152	56388	5.03	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.60%

Target Compounds					Ovalue
25) Chloroform	4.41	83	5688	0.290 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV051420\
Data File : VV016073.D
Acq On : 14 May 2020 17:44
Operator : SY/MD
Sample : L2598-16
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: May 15 08:23:14 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 15 06:27:04 2020
Response via : Initial Calibration

