

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV051420\
 Data File : VV016079.D
 Acq On : 14 May 2020 20:04
 Operator : SY/MD
 Sample : L2661-06
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H4204

Quant Time: May 15 08:56:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 15 06:27:04 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	141608	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	140687	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	63682	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	38953	5.07	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.40%
7) Chloroethane-d5	1.58	69	39040	5.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.20%
11) 1,1-Dichloroethene-d2	2.12	63	60070	3.64	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.80%
20) 2-Butanone-d5	3.95	46	121563	55.47	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.94%
24) Chloroform-d	4.38	84	83453	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.40%
26) 1,2-Dichloroethane-d4	5.06	65	48842	5.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.40%
32) Benzene-d6	5.08	84	165440	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
36) 1,2-Dichloropropane-d6	6.10	67	53161	5.20	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.00%
41) Toluene-d8	7.34	98	144575	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.40%
45) trans-1,3-Dichloropropene-	7.65	79	17388	4.34	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.80%
48) 2-Hexanone-d5	8.12	63	90933	51.61	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.22%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	37424	5.21	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.20%
65) 1,2-Dichlorobenzene-d4	11.65	152	55223	5.34	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.80%

Target Compounds

					Ovalue
13) Acetone	2.23	43	3701	2.560 ug/L	83
25) Chloroform	4.41	83	7561	0.414 ug/L	96
34) Trichloroethene	5.95	95	1566	0.155 ug/L	86
49) Tetrachloroethene	8.00	164	53531	7.594 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV051420\
Data File : VV016079.D
Acq On : 14 May 2020 20:04
Operator : SY/MD
Sample : L2661-06
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H4204

Quant Time: May 15 08:56:19 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 15 06:27:04 2020
Response via : Initial Calibration

