

Data Path : W:\HPCHEM1\MSVOA V\DATA\VV051518\
 Data File : VV005907.D
 Acq On : 15 May 2018 10:38
 Operator : SY/MD
 Sample : VV0515WBL01
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 17 13:21:42 2018
 Quant Method : W:\HPCHEM1\MSVOA V\METHOD\SOMVTR050518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 16 03:06:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	291959	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	285595	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	113281	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	119067	5.37	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	107.40%
7) Chloroethane-d5	1.59	69	91392	4.78	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.60%
11) 1,1-Dichloroethene-d2	2.13	63	176163	3.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.00%
20) 2-Butanone-d5	3.95	46	227029	52.08	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.16%
24) Chloroform-d	4.41	84	192779	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.40%
26) 1,2-Dichloroethane-d4	5.09	65	115489	4.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.40%
32) Benzene-d6	5.11	84	413152	5.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.60%
36) 1,2-Dichloropropane-d6	6.12	67	126695	5.32	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.40%
41) Toluene-d8	7.37	98	360283	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.20%
43) trans-1,3-Dichloropropene-	7.67	79	43830	4.66	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.20%
46) 2-Hexanone-d5	8.14	63	141180	38.47	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	76.94%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	74849	4.19	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	117353	4.99	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA V\DATA\VV051518\
Data File : VV005907.D
Acq On : 15 May 2018 10:38
Operator : SY/MD
Sample : VV0515WBL01
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 17 13:21:42 2018
Quant Method : W:\HPCHEM1\MSVOA V\METHOD\SOMVTR050518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 16 03:06:39 2018
Response via : Initial Calibration

