

Data Path : W:\HPCHEM1\MSVOA V\DATA\VV051518\
 Data File : VV005912.D
 Acq On : 15 May 2018 13:24
 Operator : SY/MD
 Sample : J2740-11
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 16 16:17:00 2018
 Quant Method : W:\HPCHEM1\MSVOA V\METHOD\SOMVTR050518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 16 03:06:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	254531	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	259112	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	103183	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	115957	6.00	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	120.00%
7) Chloroethane-d5	1.58	69	90183	5.41	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.20%
11) 1,1-Dichloroethene-d2	2.13	63	168564	3.95	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.00%
20) 2-Butanone-d5	3.96	46	202743	53.35	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.70%
24) Chloroform-d	4.41	84	173101	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.20%
26) 1,2-Dichloroethane-d4	5.09	65	111918	5.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.20%
32) Benzene-d6	5.11	84	393249	5.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.80%
36) 1,2-Dichloropropane-d6	6.13	67	122888	5.69	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	113.80%
41) Toluene-d8	7.37	98	339295	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.00%
43) trans-1,3-Dichloropropene-	7.67	79	40802	4.79	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.80%
46) 2-Hexanone-d5	8.14	63	144741	43.47	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.94%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	79000	4.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	117001	5.46	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA V\DATA\VV051518\
Data File : VV005912.D
Acq On : 15 May 2018 13:24
Operator : SY/MD
Sample : J2740-11
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 16 16:17:00 2018
Quant Method : W:\HPCHEM1\MSVOA V\METHOD\SOMVTR050518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 16 03:06:39 2018
Response via : Initial Calibration

