

Data Path : W:\HPCHEM1\MSVOA V\DATA\VV051518\
 Data File : VV005914.D
 Acq On : 15 May 2018 14:19
 Operator : SY/MD
 Sample : J2658-11
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 16 16:19:15 2018
 Quant Method : W:\HPCHEM1\MSVOA V\METHOD\SOMVTR050518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 16 03:06:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	271288	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	279098	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	107547	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	115445	5.60	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	112.00%
7) Chloroethane-d5	1.58	69	90640	5.10	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.00%
11) 1,1-Dichloroethene-d2	2.13	63	170551	3.75	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.00%
20) 2-Butanone-d5	3.96	46	229744	56.72	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.44%
24) Chloroform-d	4.41	84	183097	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.60%
26) 1,2-Dichloroethane-d4	5.09	65	113357	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.80%
32) Benzene-d6	5.11	84	397659	5.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.00%
36) 1,2-Dichloropropane-d6	6.13	67	120710	5.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.80%
41) Toluene-d8	7.37	98	348638	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.40%
43) trans-1,3-Dichloropropene-	7.67	79	42862	4.67	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.40%
46) 2-Hexanone-d5	8.14	63	148870	41.51	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.02%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	79002	4.52	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	115559	5.17	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA V\DATA\VV051518\
Data File : VV005914.D
Acq On : 15 May 2018 14:19
Operator : SY/MD
Sample : J2658-11
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 16 16:19:15 2018
Quant Method : W:\HPCHEM1\MSVOA V\METHOD\SOMVTR050518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 16 03:06:39 2018
Response via : Initial Calibration

