

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW051723\
 Data File : VW031081.D
 Acq On : 17 May 2023 18:10
 Operator : SY/MD
 Sample : 02822-10
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
ClientSampleId :
 CONN8

Quant Time: May 18 02:52:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu May 18 02:47:29 2023
 Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|------------------------------------|--------|----------------|----------|--------|----------|----------|
| Internal Standards | | | | | | |
| 1) 1,4-Difluorobenzene | 5.542 | 114 | 121922 | 5.000 | ug/L | 0.00 |
| 28) Chlorobenzene-d5 | 8.793 | 117 | 122151 | 5.000 | ug/L | 0.00 |
| 58) 1,4-Dichlorobenzene-d4 | 11.194 | 152 | 63699 | 5.000 | ug/L | 0.00 |
| System Monitoring Compounds | | | | | | |
| 4) Vinyl Chloride-d3 | 1.282 | 65 | 32303 | 3.216 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 40 - 130 | Recovery | = | 64.400% | |
| 7) Chloroethane-d5 | 1.539 | 69 | 35154 | 4.290 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 65 - 130 | Recovery | = | 85.800% | |
| 11) 1,1-Dichloroethene-d2 | 2.063 | 65 | 18264 | 3.436 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 60 - 125 | Recovery | = | 68.800% | |
| 20) 2-Butanone-d5 | 3.818 | 46 | 108894 | 61.132 | ug/L | -0.02 |
| Spiked Amount | 50.000 | Range 40 - 130 | Recovery | = | 122.260% | |
| 24) Chloroform-d | 4.256 | 84 | 98110 | 4.915 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 70 - 125 | Recovery | = | 98.200% | |
| 26) 1,2-Dichloroethane-d4 | 4.953 | 65 | 52537 | 5.177 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 70 - 130 | Recovery | = | 103.600% | |
| 32) Benzene-d6 | 4.970 | 84 | 155981 | 4.356 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 70 - 125 | Recovery | = | 87.200% | |
| 36) 1,2-Dichloropropane-d6 | 5.995 | 67 | 49531 | 4.757 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 60 - 140 | Recovery | = | 95.200% | |
| 41) Toluene-d8 | 7.252 | 98 | 131157 | 3.847 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 70 - 130 | Recovery | = | 77.000% | |
| 43) trans-1,3-Dichloroprop... | 7.561 | 79 | 16098 | 4.125 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 55 - 130 | Recovery | = | 82.400% | |
| 46) 2-Hexanone-d5 | 8.034 | 63 | 76447 | 55.908 | ug/L | 0.00 |
| Spiked Amount | 50.000 | Range 45 - 130 | Recovery | = | 111.820% | |
| 56) 1,1,2,2-Tetrachloroeth... | 10.159 | 84 | 38812 | 5.143 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 65 - 120 | Recovery | = | 102.800% | |
| 66) 1,2-Dichlorobenzene-d4 | 11.567 | 152 | 57233 | 4.820 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 80 - 120 | Recovery | = | 96.400% | |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051723\
 Data File : VV031081.D
 Acq On : 17 May 2023 18:10
 Operator : SY/MD
 Sample : 02822-10
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CONN8

Quant Time: May 18 02:52:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu May 18 02:47:29 2023
 Response via : Initial Calibration

