

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW051723\
 Data File : VW031082.D
 Acq On : 17 May 2023 18:34
 Operator : SY/MD
 Sample : 02822-11
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CONP7

Quant Time: May 18 02:52:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu May 18 02:47:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	126778	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	123565	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	65304	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	29094	2.785	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	55.800%
7) Chloroethane-d5	1.539	69	32915	3.863	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.200%
11) 1,1-Dichloroethene-d2	2.063	65	16544	2.993	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	59.800%#
20) 2-Butanone-d5	3.815	46	102061	55.101	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.200%
24) Chloroform-d	4.259	84	87673	4.224	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.400%
26) 1,2-Dichloroethane-d4	4.953	65	48813	4.625	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.600%
32) Benzene-d6	4.970	84	141025	3.894	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.800%
36) 1,2-Dichloropropane-d6	5.995	67	44362	4.212	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.200%
41) Toluene-d8	7.249	98	118938	3.448	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	69.000%#
43) trans-1,3-Dichloroprop...	7.561	79	15421	3.906	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.200%
46) 2-Hexanone-d5	8.034	63	68651	49.632	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.260%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	36592	4.793	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.800%
66) 1,2-Dichlorobenzene-d4	11.567	152	53173	4.368	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051723\
Data File : VV031082.D
Acq On : 17 May 2023 18:34
Operator : SY/MD
Sample : 02822-11
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CONP7

Quant Time: May 18 02:52:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu May 18 02:47:29 2023
Response via : Initial Calibration

