

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW051723\
 Data File : VW031069.D
 Acq On : 17 May 2023 13:23
 Operator : SY/MD
 Sample : 02822-14
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CONQ1

Quant Time: May 18 02:49:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu May 18 02:47:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	127948	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	126494	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	63105	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	31023	2.943	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	58.800%
7) Chloroethane-d5	1.539	69	34010	3.955	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.000%
11) 1,1-Dichloroethene-d2	2.063	65	18178	3.258	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.200%
20) 2-Butanone-d5	3.825	46	84671	45.295	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.580%
24) Chloroform-d	4.259	84	90732	4.331	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.600%
26) 1,2-Dichloroethane-d4	4.950	65	48032	4.510	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.200%
32) Benzene-d6	4.969	84	148637	4.009	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.200%
36) 1,2-Dichloropropane-d6	5.998	67	44337	4.112	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	82.200%
41) Toluene-d8	7.249	98	122931	3.482	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	69.600%#
43) trans-1,3-Dichloroprop...	7.564	79	13353	3.304	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	66.000%
46) 2-Hexanone-d5	8.034	63	59548	42.054	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.100%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	33759	4.320	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.400%
66) 1,2-Dichlorobenzene-d4	11.567	152	52582	4.470	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051723\
Data File : VV031069.D
Acq On : 17 May 2023 13:23
Operator : SY/MD
Sample : 02822-14
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CONQ1

Quant Time: May 18 02:49:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu May 18 02:47:29 2023
Response via : Initial Calibration

