

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW051723\
 Data File : VW031070.D
 Acq On : 17 May 2023 13:45
 Operator : SY/MD
 Sample : 02822-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CONK5

Quant Time: May 18 02:49:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu May 18 02:47:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	120845	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	121032	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	61951	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	30122	3.025	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	60.600%
7) Chloroethane-d5	1.539	69	33171	4.084	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.600%
11) 1,1-Dichloroethene-d2	2.066	65	16361	3.105	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.200%
20) 2-Butanone-d5	3.818	46	86216	48.832	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.660%
24) Chloroform-d	4.262	84	89709	4.534	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.600%
26) 1,2-Dichloroethane-d4	4.953	65	47249	4.697	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.000%
32) Benzene-d6	4.969	84	145537	4.102	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.000%
36) 1,2-Dichloropropane-d6	5.998	67	44130	4.278	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.600%
41) Toluene-d8	7.252	98	120229	3.559	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.200%
43) trans-1,3-Dichloroprop...	7.564	79	14549	3.762	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.200%
46) 2-Hexanone-d5	8.034	63	61357	45.287	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.580%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	34099	4.560	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.200%
66) 1,2-Dichlorobenzene-d4	11.567	152	50974	4.414	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051723\
Data File : VV031070.D
Acq On : 17 May 2023 13:45
Operator : SY/MD
Sample : 02822-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CONK5

Quant Time: May 18 02:49:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu May 18 02:47:29 2023
Response via : Initial Calibration

