

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051822\
 Data File : VV025928.D
 Acq On : 18 May 2022 13:26
 Operator : SY/MD
 Sample : VSTD0.102
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.1202

Quant Time: May 19 07:38:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM051822.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 19 07:35:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	21124	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	18697	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	10802	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	1691	0.114	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	978	0.112	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	1172	0.116	ug/L	0.00
8) Toluene-d8	7.317	98	4377	0.113	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	894	0.114	ug/L	0.00
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.310	62	1382	0.105	ug/L #	65
6) Trichloroethene	5.925	95	1397	0.108	ug/L	94
9) 1,2-Dibromoethane	8.368	107	832	0.105	ug/L #	94
12) 1,2,3-Trichloropropane	10.267	75	506	0.104	ug/L	95
13) 1,2-Dibromo-3-chloropr...	12.426	75	110	0.128	ug/L #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051822\
Data File : VV025928.D
Acq On : 18 May 2022 13:26
Operator : SY/MD
Sample : VSTD0.102
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.1202

Quant Time: May 19 07:38:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM051822.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 19 07:35:53 2022
Response via : Initial Calibration

