

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051822\
 Data File : VV025927.D
 Acq On : 18 May 2022 12:53
 Operator : SY/MD
 Sample : VSTD0.0501
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.05201

Quant Time: May 19 07:38:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM051822.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 19 07:35:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	16055	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	14971	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	8747	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	641	0.057	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.066	65	356	0.054	ug/L	0.02
7) 1,2-Dichloropropane-d6	6.092	67	435	0.054	ug/L	0.02
8) Toluene-d8	7.332	98	1640	0.053	ug/L	0.02
10) 1,1,2,2-Tetrachloroeth...	10.216	84	333	0.053	ug/L	0.00
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.310	62	555	0.056	ug/L #	1
6) Trichloroethene	5.925	95	570	0.055	ug/L	83
9) 1,2-Dibromoethane	8.368	107	359	0.056	ug/L #	92
12) 1,2,3-Trichloropropane	10.267	75	214	0.055	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.426	75	32	0.046	ug/L #	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051822\
Data File : VV025927.D
Acq On : 18 May 2022 12:53
Operator : SY/MD
Sample : VSTD0.0501
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.05201

Quant Time: May 19 07:38:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM051822.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 19 07:35:53 2022
Response via : Initial Calibration

