

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051823\
 Data File : VV031092.D
 Acq On : 18 May 2023 15:32
 Operator : SY/MD
 Sample : 02822-15
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: May 19 05:36:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri May 19 05:32:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	115004	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	112240	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	56113	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	29516	3.115	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.400%
7) Chloroethane-d5	1.539	69	32392	4.191	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.800%
11) 1,1-Dichloroethene-d2	2.063	65	16732	3.337	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.800%
20) 2-Butanone-d5	3.818	46	87360	51.993	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.980%
24) Chloroform-d	4.259	84	88936	4.723	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.400%
26) 1,2-Dichloroethane-d4	4.950	65	47402	4.952	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.000%
32) Benzene-d6	4.966	84	142388	4.328	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.600%
36) 1,2-Dichloropropane-d6	5.995	67	45928	4.801	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.000%
41) Toluene-d8	7.249	98	116841	3.729	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.600%
43) trans-1,3-Dichloroprop...	7.561	79	14939	4.166	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.400%
46) 2-Hexanone-d5	8.034	63	59908	47.682	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.360%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	35239	5.082	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.600%
66) 1,2-Dichlorobenzene-d4	11.567	152	48861	4.671	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.400%
Target Compounds						Qvalue
16) Methylene chloride	2.458	84	2949	0.337	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051823\
Data File : VV031092.D
Acq On : 18 May 2023 15:32
Operator : SY/MD
Sample : 02822-15
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: May 19 05:36:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri May 19 05:32:32 2023
Response via : Initial Calibration

